

Asset Management Back Save Apply Edit Asset Copy Asset Verification

Description
Cardiac Monitor

Asset Identification
 Division:* Main Division
 Asset Tag Number : A00806
 Serial Number :* AR13L006405
 Internal Serial Number:

Categorization
 Asset Type:* Fixed Asset
 Asset Class:* Durable Medical Equipment
 Category:* Fixed Asset
 Subcategory:* Cardiac Monitor
 General Ledger Number: Select General Ledger
 Status: ENABLED

Notes

Scheduled Maintenance Add			
Maintenance	Type	Schedule	Maintenance Due ^A
Preventative Maintenance	Variable	Every 12 months	04/07/2026

Asset Management
Asset Management
 Model Year: 2013
 Model Number: X-S
 Manufacturer: Zoll
 Purchase Date: 12/09/2013
 In Service Date: 12/09/2013
 Warranty Expiration Date:
 Planned Decommission
 Date:
 Decommission Date:
 Asset Value :* 23300.000

Fixed Asset Last Seen Location History			
Location	Date	Crew Member	Source

Supply Room Availability					
Supply Room	Stock Location	On Hand	Verification Date	Verified By	Status
DME - Capital Assets	Select Stock Location	1 EA			Not Verified

Asset Check Out Check In History				
<input type="checkbox"/>	Check Out To	Check Out Date	Due Back	Verification Schedule StatNext Verification D:
No CheckOuts.				

Historical Maintenance Records New Print Delete							
<input type="checkbox"/>	WO #	Date	Status	Type	Description	Cost	Attachments
<input type="checkbox"/>	2320	04/07/2025	In-Process	Preventative Maintenance	Annual PM by Zoll Tech, Tom Gillis	\$0.00	View
<input type="checkbox"/>	2253	04/04/2024	Closed	Preventative Maintenance	Annual PM by Tom Gillis, Zoll	\$0.00	View
<input type="checkbox"/>	2179	04/26/2023	Closed	Preventative Maintenance	Annual PM by tom Gillis	\$0.00	View
<input type="checkbox"/>	2111	08/23/2022	Closed	Repair Work Order	Defib Disabled- Replaced PACE Board	\$0.00	View
<input type="checkbox"/>	2110	08/22/2022	Closed	Repair Work Order	Power Port had damaged pins and they were repaired	\$0.00	View